

Digital mammography cancer detection rates may vary significantly

May 14 2013

Digital direct radiography (DR) is significantly more effective than computed radiography (CR) at detecting breast cancer, according to a new study published online in the journal *Radiology*.

The findings are the first to show a difference between the two types of digital imaging when compared with screen film mammography and suggest that women should be informed of the potential for lower [cancer detection](#) with CR, the researchers said.

Digital mammography, which takes an electronic image of the breast that can be stored and sent electronically, has supplanted screen film mammography in recent years.

DR and CR are the two types of digital mammography offered. DR is an online system in which the detector is an integral part of the mammographic unit, and the [digital image](#) can be read by the system in real time. CR is an offline system that relies on a cassette-based removable detector. An external [reading device](#) is used to generate the digital image.

While numerous studies have been published comparing digital mammography with screen film mammography, less is known about the [comparative effectiveness](#) of DR and CR, according to Anna M. Chiarelli, Ph.D., senior scientist in Prevention and [Cancer Control](#) at Cancer Care Ontario in Toronto, Canada.

"[Digital mammography](#) was implemented progressively in Ontario starting in 2006, and since that time there's been no evidence in the literature to suggest that CR did not perform as well as DR," she said.

For the new study, Dr. Chiarelli and colleagues drew upon information from the Ontario Breast Screening Program (OBSP), initiated in 1990 to deliver breast screening to women between the ages of 50 and 74. The researchers identified three groups of women 50 to 74 years old who were screened between Jan. 1, 2008, and Dec. 31, 2009. A total of 403,688 women were screened by screen film mammography, while 220,520 had DR and 64,210 underwent CR. The women were followed for 12 months after screening.

DR detected 4.9 cancers per 1,000 mammograms, a figure almost identical to screen film mammography's rate of 4.8 cancers per 1,000 mammograms. However, CR's detection rate of 3.4 cancers per 1,000 mammograms was significantly lower.

"CR was 21 percent less effective than DR," Dr. Chiarelli said. "This could result in about 10 fewer cancers detected per 10,000 women screened."

The reason for the lower detection rates likely can be attributed to technical factors, Dr. Chiarelli added.

"There may be several technical reasons reported by others for the lower effectiveness of CR, including loss of spatial resolution, or sharpness, and increased image noise, or granularity," she said.

Dr. Chiarelli said the new results carry more weight than those of previous studies because of the larger number of women screened and the reliance on concurrent cohorts, or different groups followed over the same time period.

More information: "Digital Compared with Screen-Film Mammography: Performance Measures in Concurrent Cohorts within an Organized Breast Screening Program" *Radiology*, 2013.

Provided by Radiological Society of North America

Citation: Digital mammography cancer detection rates may vary significantly (2013, May 14)
retrieved 3 February 2024 from

<https://medicalxpress.com/news/2013-05-digital-mammography-cancer-vary-significantly.html>

This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.